

General purpose power unit

P704B

SPECIFICATIONS

INPUT CHARACTERISTICS

Voltage to transducer	27 VDC ¹
Current to transducer, $\pm 20\%$	2.4 mA DC

OUTPUT CHARACTERISTICS

Output impedance	same as transducer
Recommended load impedance	>100 k Ω
Decoupling capacitor	22 μ F, 35 VDC

TRANSFER CHARACTERISTICS

Frequency response	same as transducer
--------------------	--------------------

BATTERY TEST CIRCUIT / POWER REQUIREMENTS

LED lights	>18 VDC
Battery life	>120 hours
Batteries ²	9V alkaline (3)

ENVIRONMENTAL

Temperature range	0° to +55°C
-------------------	-------------

PHYSICAL CHARACTERISTICS

Size (W x H x D)	3 x 2.4 x 4 inches
Weight	0.84 lb
Connectors:	
Signal input	BNC
Signal output	BNC

Notes: ¹ 25 VDC when using Ni-Cad batteries.

² For extended operation, the NC3 Ni-Cad battery kit should be used.

Accessories supplied: (3) 9V alkaline batteries

Accessories available: NC3 Ni-Cad battery kit; LA704B line adaptor (110V); LA704B-220 line adaptor (220V); BNC series adaptors



Key features

- Powers most 700 series accelerometers
- Portable and lightweight
- Battery condition light
- Manufactured in an ISO 9001 facility

See reverse for operating instructions.


Note: Due to continuous process improvement, specifications are subject to change without notice.
This document is cleared for public release.

General purpose power unit P704B

Operating instructions

1. To test batteries, press switch to RIGHT (momentary) position; LED should light.
2. OFF position is with switch in center.
3. To turn unit ON, press to LEFT.


Use LA704B line adaptor to power unit from line voltage without batteries installed or to charge Ni-Cad batteries.



**CAUTION: DO NOT ATTEMPT TO RECHARGE ALKALINES WITH THE LA704B.
Alkaline batteries may EXPLODE or leak corrosive fluids.**

Test for proper operation

- Use a digital multimeter to verify that the proper voltage and current are available at the transducer connector.
- Substitute an oscillator for the transducer.
- Follow the connection to the power unit as shown at right.
- The unit should have unity again.



Note: Due to continuous process improvement, specifications are subject to change without notice.
This document is cleared for public release.



Компания «ЭлектроПласт» предлагает заключение долгосрочных отношений при поставках импортных электронных компонентов на взаимовыгодных условиях!

Наши преимущества:

- Оперативные поставки широкого спектра электронных компонентов отечественного и импортного производства напрямую от производителей и с крупнейших мировых складов;
- Поставка более 17-ти миллионов наименований электронных компонентов;
- Поставка сложных, дефицитных, либо снятых с производства позиций;
- Оперативные сроки поставки под заказ (от 5 рабочих дней);
- Экспресс доставка в любую точку России;
- Техническая поддержка проекта, помощь в подборе аналогов, поставка прототипов;
- Система менеджмента качества сертифицирована по Международному стандарту ISO 9001;
- Лицензия ФСБ на осуществление работ с использованием сведений, составляющих государственную тайну;
- Поставка специализированных компонентов (Xilinx, Altera, Analog Devices, Intersil, Interpoint, Microsemi, Aeroflex, Peregrine, Syfer, Eurofarad, Texas Instrument, Miteq, Cobham, E2V, MA-COM, Hittite, Mini-Circuits, General Dynamics и др.);

Помимо этого, одним из направлений компании «ЭлектроПласт» является направление «Источники питания». Мы предлагаем Вам помощь Конструкторского отдела:

- Подбор оптимального решения, техническое обоснование при выборе компонента;
- Подбор аналогов;
- Консультации по применению компонента;
- Поставка образцов и прототипов;
- Техническая поддержка проекта;
- Защита от снятия компонента с производства.



Как с нами связаться

Телефон: 8 (812) 309 58 32 (многоканальный)

Факс: 8 (812) 320-02-42

Электронная почта: org@eplast1.ru

Адрес: 198099, г. Санкт-Петербург, ул. Калинина, дом 2, корпус 4, литера А.